

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080919\
 Data File : VV012149.D
 Acq On : 09 Aug 2019 12:36
 Operator : SY/MD
 Sample : K4184-01
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA0

Quant Time: Aug 10 03:08:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 10 02:29:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	135587	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	124764	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	66899	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22924	3.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.40%
7) Chloroethane-d5	1.58	69	17363	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.00%
11) 1,1-Dichloroethene-d2	2.13	63	35581	3.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.00%
20) 2-Butanone-d5	3.96	46	58195	39.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.16%
24) Chloroform-d	4.40	84	65321	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.00%
26) 1,2-Dichloroethane-d4	5.08	65	34929	4.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.00%
32) Benzene-d6	5.09	84	127738	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
36) 1,2-Dichloropropane-d6	6.11	67	32799	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.00%
41) Toluene-d8	7.36	98	124088	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.80%
43) trans-1,3-Dichloropropene-	7.66	79	15498	3.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.20%
46) 2-Hexanone-d5	8.13	63	48559	36.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	26540	3.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	52706	4.24	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.80%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	1243	0.137 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080919\
Data File : VV012149.D
Acq On : 09 Aug 2019 12:36
Operator : SY/MD
Sample : K4184-01
Misc : 25mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA0

Quant Time: Aug 10 03:08:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 10 02:29:42 2019
Response via : Initial Calibration

