

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080921\
 Data File : VV021759.D
 Acq On : 09 Aug 2021 11:35
 Operator : SY/MD
 Sample : VV0809WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK229

Quant Time: Aug 10 01:21:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 10 01:21:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	388655	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	363481	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	175307	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	103235	42.296	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.600%
7) Chloroethane-d5	1.564	69	82916	41.189	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.380%
11) 1,1-Dichloroethene-d2	2.108	63	134423	30.793	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.580%
21) 2-Butanone-d5	3.879	46	218149	104.048	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.050%
24) Chloroform-d	4.349	84	245250	47.406	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.820%
26) 1,2-Dichloroethane-d4	5.034	65	152511	49.295	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.580%
32) Benzene-d6	5.050	84	494527	48.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	6.072	67	162734	49.837	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.680%
41) Toluene-d8	7.317	98	447863	46.986	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.980%
43) trans-1,3-Dichloroprop...	7.625	79	75718	46.994	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.980%
47) 2-Hexanone-d5	8.092	63	134876	94.728	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.730%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	198307	46.081	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.160%
66) 1,2-Dichlorobenzene-d4	11.628	152	180758	49.800	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.600%
Target Compounds					Qvalue	
48) 2-Hexanone	8.146	43	17005	5.847	ug/L	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080921\
Data File : VV021759.D
Acq On : 09 Aug 2021 11:35
Operator : SY/MD
Sample : VV0809WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK229

Quant Time: Aug 10 01:21:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080221WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 10 01:21:22 2021
Response via : Initial Calibration

