

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081018\
 Data File : VV006986.D
 Acq On : 10 Aug 2018 16:10
 Operator : SY/MD
 Sample : J4422-09
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD1

Quant Time: Aug 11 00:12:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 10 23:13:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	244707	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	233737	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	94836	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	67054	4.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.00%
7) Chloroethane-d5	1.58	69	60823	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.80%
11) 1,1-Dichloroethene-d2	2.13	63	90028	3.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	3.97	46	227922	50.17	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.34%
24) Chloroform-d	4.41	84	152975	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.09	65	79865	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.10	84	284583	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
36) 1,2-Dichloropropane-d6	6.12	67	99896	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.37	98	201924	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
43) trans-1,3-Dichloropropene-	7.67	79	30334	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	8.15	63	111829	38.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.58%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	74971	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	79319	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%

Target Compounds					Ovalue
13) Acetone	2.21	43	9990	4.27 ug/L	96
25) Chloroform	4.43	83	13219	0.36 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081018\
Data File : VV006986.D
Acq On : 10 Aug 2018 16:10
Operator : SY/MD
Sample : J4422-09
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD1

Quant Time: Aug 11 00:12:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 10 23:13:31 2018
Response via : Initial Calibration

