

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081018\
 Data File : VV006989.D
 Acq On : 10 Aug 2018 17:39
 Operator : SY/MD
 Sample : VV0810WBL02
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK85

Quant Time: Aug 11 00:24:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 11 00:19:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	248738	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	237704	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	95697	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	69196	4.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.20%
7) Chloroethane-d5	1.58	69	63667	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.40%
11) 1,1-Dichloroethene-d2	2.13	63	95036	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	3.97	46	238734	51.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.40%
24) Chloroform-d	4.40	84	159328	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.09	65	83814	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	5.10	84	296327	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
36) 1,2-Dichloropropane-d6	6.12	67	104783	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.36	98	219331	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	7.67	79	31761	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
46) 2-Hexanone-d5	8.15	63	126533	42.60	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.20%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	77215	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	83892	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081018\
Data File : VV006989.D
Acq On : 10 Aug 2018 17:39
Operator : SY/MD
Sample : VV0810WBL02
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK85

Quant Time: Aug 11 00:24:43 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 11 00:19:35 2018
Response via : Initial Calibration

