

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081019\
 Data File : VV012163.D
 Acq On : 09 Aug 2019 21:11
 Operator : SY/MD
 Sample : VV0810WBL01
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK63

Quant Time: Aug 12 04:04:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081019WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 05:07:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	289831	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	260056	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	117903	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50188	47.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.14%
7) Chloroethane-d5	1.58	69	33420	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.32%
11) 1,1-Dichloroethene-d2	2.13	63	69233	34.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.70%
21) 2-Butanone-d5	3.93	46	40840	77.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.47%
24) Chloroform-d	4.40	84	98109	44.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.58%
26) 1,2-Dichloroethane-d4	5.08	65	63621	50.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.16%
32) Benzene-d6	5.10	84	206512	51.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.44%
36) 1,2-Dichloropropane-d6	6.12	67	55801	50.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.76%
41) Toluene-d8	7.36	98	186174	48.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.94%
43) trans-1,3-Dichloropropene-	7.66	79	27557	46.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.72%
47) 2-Hexanone-d5	8.13	63	40226	95.87	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.87%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	73314	48.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.52%
64) 1,2-Dichlorobenzene-d4	11.67	152	71200	51.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081019\
Data File : VV012163.D
Acq On : 09 Aug 2019 21:11
Operator : SY/MD
Sample : VV0810WBL01
Misc : 5mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK63

Quant Time: Aug 12 04:04:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081019WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 05:07:48 2019
Response via : Initial Calibration

