

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081020\
 Data File : VV017888.D
 Acq On : 10 Aug 2020 11:14
 Operator : SY/MD
 Sample : L3608-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 11 02:48:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 11 01:17:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	226231	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	212115	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	109341	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	56560	48.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.82%
7) Chloroethane-d5	1.58	69	48280	48.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.28%
11) 1,1-Dichloroethene-d2	2.12	63	76960m	27.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	54.84%#
21) 2-Butanone-d5	3.91	46	73068	99.88	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.88%
24) Chloroform-d	4.37	84	147339	48.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.20%
26) 1,2-Dichloroethane-d4	5.06	65	114839	51.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.20%
32) Benzene-d6	5.08	84	252978	51.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.30%
36) 1,2-Dichloropropane-d6	6.09	67	64412	51.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.72%
41) Toluene-d8	7.34	98	248476	50.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.94%
43) trans-1,3-Dichloropropene-	7.64	79	25224	31.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.08%
47) 2-Hexanone-d5	8.11	63	45912	102.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.07%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	97651	50.44	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.88%
64) 1,2-Dichlorobenzene-d4	11.65	152	117049	52.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.02%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	22100	16.719 ug/L	94
14) Carbon disulfide	2.31	76	3027	0.985 ug/L #	81
17) trans-1,2-Dichloroethene	2.78	96	4230	3.110 ug/L	99
20) cis-1,2-Dichloroethene	3.94	96	123018	83.731 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081020\
Data File : VV017888.D
Acq On : 10 Aug 2020 11:14
Operator : SY/MD
Sample : L3608-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 11 02:48:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 01:17:50 2020
Response via : Initial Calibration

