

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081020\
 Data File : VV017903.D
 Acq On : 10 Aug 2020 18:03
 Operator : SY/MD
 Sample : L3608-07DL 5X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F2L78DL

Quant Time: Aug 11 03:25:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 11 01:17:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	245205	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	229495	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	117545	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	48840	38.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.14%
7) Chloroethane-d5	1.58	69	44316	41.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.38%
11) 1,1-Dichloroethene-d2	2.12	63	92206	30.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.60%
21) 2-Butanone-d5	3.91	46	79877	100.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.73%
24) Chloroform-d	4.38	84	146722	44.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.38%
26) 1,2-Dichloroethane-d4	5.05	65	110562	45.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.66%
32) Benzene-d6	5.07	84	242482	45.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.50%
36) 1,2-Dichloropropane-d6	6.09	67	64181	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.52%
41) Toluene-d8	7.34	98	236136	44.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.66%
43) trans-1,3-Dichloropropene-	7.64	79	40771	47.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.24%
47) 2-Hexanone-d5	8.11	63	47247	97.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.08%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	96583	46.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	115234	48.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.18%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	43859	30.613 ug/L	93
17) trans-1,2-Dichloroethene	2.78	96	1853	1.257 ug/L #	76
20) cis-1,2-Dichloroethene	3.94	96	103980	65.296 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081020\
Data File : VV017903.D
Acq On : 10 Aug 2020 18:03
Operator : SY/MD
Sample : L3608-07DL 5X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F2L78DL

Quant Time: Aug 11 03:25:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 01:17:50 2020
Response via : Initial Calibration

