

Data File : VV017886.D
Acq On : 10 Aug 2020 10:27
Operator : SY/MD
Sample : VV0810WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Aug 11 01:27:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 01:17:50 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	249881	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	235100	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	123864	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54394	42.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.30%
7) Chloroethane-d5	1.58	69	49312	44.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.96%
11) 1,1-Dichloroethene-d2	2.12	63	104081	33.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.14%
21) 2-Butanone-d5	3.91	46	82953	102.66	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.66%
24) Chloroform-d	4.38	84	156783	46.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.68%
26) 1,2-Dichloroethane-d4	5.06	65	120604	49.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.12%
32) Benzene-d6	5.08	84	263361	48.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.02%
36) 1,2-Dichloropropane-d6	6.10	67	69149	50.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.46%
41) Toluene-d8	7.34	98	263816	48.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.70%
43) trans-1,3-Dichloropropene-	7.64	79	45398	51.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.44%
47) 2-Hexanone-d5	8.11	63	48836	97.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.96%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	101079	47.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	124638	49.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.72%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV017886.D
Acq On : 10 Aug 2020 10:27
Operator : SY/MD
Sample : VV0810WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Aug 11 01:27:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 01:17:50 2020
Response via : Initial Calibration

