

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081022\
 Data File : VW027269.D
 Acq On : 10 Aug 2022 20:00
 Operator : SY/MD
 Sample : N4144-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY144

Quant Time: Aug 11 03:55:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 09 02:29:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	314781	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	328220	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	134573	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	110827	51.641	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 103.280%	
7) Chloroethane-d5	1.558	69	96817	61.709	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 123.420%	
11) 1,1-Dichloroethene-d2	2.101	63	155977	39.734	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 79.460%	
21) 2-Butanone-d5	3.870	46	227915	119.308	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 119.310%	
24) Chloroform-d	4.342	84	240919	54.803	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.600%	
26) 1,2-Dichloroethane-d4	5.027	65	163833	60.025	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 120.060%	
32) Benzene-d6	5.043	84	469131	59.603	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 119.200%	
36) 1,2-Dichloropropane-d6	6.066	67	169793	60.644	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 121.280%#	
41) Toluene-d8	7.313	98	383935	57.227	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 114.460%	
43) trans-1,3-Dichloroprop...	7.622	79	65344	53.544	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 107.080%	
47) 2-Hexanone-d5	8.085	63	151375	112.254	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 112.250%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	250638	54.932	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 109.860%	
66) 1,2-Dichlorobenzene-d4	11.622	152	151574	65.223	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 130.440%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081022\
Data File : VV027269.D
Acq On : 10 Aug 2022 20:00
Operator : SY/MD
Sample : N4144-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY144

Quant Time: Aug 11 03:55:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 09 02:29:12 2022
Response via : Initial Calibration

