

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081022\
 Data File : VW027270.D
 Acq On : 10 Aug 2022 20:25
 Operator : SY/MD
 Sample : N4144-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY147

Quant Time: Aug 11 03:55:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 09 02:29:12 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	322093	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	340387	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	142317	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	114192	52.001	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	104.000%	
7) Chloroethane-d5	1.558	69	99927	62.245	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	124.500%	
11) 1,1-Dichloroethene-d2	2.101	63	158552	39.473	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	78.940%	
21) 2-Butanone-d5	3.870	46	237255	121.378	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	121.380%	
24) Chloroform-d	4.342	84	252148	56.056	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	112.120%	
26) 1,2-Dichloroethane-d4	5.027	65	168477	60.325	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	120.660%	
32) Benzene-d6	5.043	84	481206	58.952	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	117.900%	
36) 1,2-Dichloropropane-d6	6.066	67	173790	59.853	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	119.700%	
41) Toluene-d8	7.313	98	399666	57.442	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	114.880%	
43) trans-1,3-Dichloroprop...	7.619	79	65719	51.926	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	103.860%	
47) 2-Hexanone-d5	8.085	63	154416	110.416	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	110.420%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	252057	53.268	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	106.540%	
66) 1,2-Dichlorobenzene-d4	11.625	152	153846	62.598	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	125.200%#	
Target Compounds						
					Qvalue	
13) Acetone	2.159	43	24213	16.065	ug/L	83
19) 1,1-Dichloroethane	3.185	63	25437	5.131	ug/L	99
20) cis-1,2-Dichloroethene	3.908	96	10249	3.995	ug/L	91
30) 1,1,1-Trichloroethane	4.600	97	14279	3.165	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081022\
Data File : VV027270.D
Acq On : 10 Aug 2022 20:25
Operator : SY/MD
Sample : N4144-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY147

Quant Time: Aug 11 03:55:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 09 02:29:12 2022
Response via : Initial Calibration





#13

Acetone

Concen: 16.065 ug/L

RT: 2.159 min Scan# 34

Delta R.T. 0.006 min

Lab File: VV027270.D

Acq: 10 Aug 2022 20:25

Instrument :

MSVOA_V

ClientSampleId :





EY147

Tgt Ion: 43 Resp: 24213

Ion Ratio Lower Upper

43 100

58 33.2 0.0 49.6



#19

1,1-Dichloroethane

Concen: 5.131 ug/L

RT: 3.185 min Scan# 667

Delta R.T. 0.006 min

Lab File: VV027270.D

Acq: 10 Aug 2022 20:25




Tgt Ion: 63 Resp: 25437


Ion Ratio Lower Upper

63 100

65 32.1 22.7 42.3

83 13.7 9.5 17.7









#20
 cis-1,2-Dichloroethene
 Concen: 3.995 ug/L
 RT: 3.908 min Scan# 890
 Delta R.T. 0.009 min
 Lab File: VV027270.D
 Acq: 10 Aug 2022 20:25

Instrument :
 MSVOA_V
 ClientSampleId :
 EY147

Tgt Ion: 96 Resp: 10249
 Ion Ratio Lower Upper
 96 100
 61 126.2 81.8 151.8
 68 0.0 0.0 0.0



#30
 1,1,1-Trichloroethane
 Concen: 3.165 ug/L
 RT: 4.600 min Scan# 1107
 Delta R.T. 0.003 min
 Lab File: VV027270.D
 Acq: 10 Aug 2022 20:25

Tgt Ion: 97 Resp: 14279
 Ion Ratio Lower Upper
 97 100
 99 60.7 51.4 77.2
 61 44.8 32.1 48.1

