

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV081121\  
 Data File : VV021770.D  
 Acq On : 11 Aug 2021 12:58  
 Operator : SY/MD  
 Sample : VV0811WBL01  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK207

Quant Time: Aug 12 03:13:25 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081121WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Aug 12 03:04:55 2021  
 Response via : Initial Calibration

| Compound                      | R.T. QIon |                | Response | Conc       | Units  | Dev(Min) |
|-------------------------------|-----------|----------------|----------|------------|--------|----------|
| -----                         |           |                |          |            |        |          |
| Internal Standards            |           |                |          |            |        |          |
| 1) 1,4-Difluorobenzene        | 5.619     | 114            | 195460   | 5.000      | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 8.854     | 117            | 183511   | 5.000      | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.252    | 152            | 90569    | 5.000      | ug/L   | 0.00     |
| System Monitoring Compounds   |           |                |          |            |        |          |
| 4) Vinyl Chloride-d3          | 1.304     | 65             | 39969    | 5.341      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 40 - 130 | Recovery | = 106.800% |        |          |
| 7) Chloroethane-d5            | 1.564     | 69             | 48264    | 5.451      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 65 - 130 | Recovery | = 109.000% |        |          |
| 11) 1,1-Dichloroethene-d2     | 2.101     | 63             | 64755    | 3.816      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 60 - 125 | Recovery | = 76.400%  |        |          |
| 20) 2-Butanone-d5             | 3.918     | 46             | 148375   | 49.880     | ug/L   | 0.00     |
| Spiked Amount                 | 50.000    | Range 40 - 130 | Recovery | = 99.760%  |        |          |
| 24) Chloroform-d              | 4.349     | 84             | 126594   | 5.357      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 70 - 125 | Recovery | = 107.200% |        |          |
| 26) 1,2-Dichloroethane-d4     | 5.034     | 65             | 59912    | 5.542      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 70 - 130 | Recovery | = 110.800% |        |          |
| 32) Benzene-d6                | 5.050     | 84             | 259680   | 5.673      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 70 - 125 | Recovery | = 113.400% |        |          |
| 36) 1,2-Dichloropropane-d6    | 6.072     | 67             | 81368    | 5.511      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 60 - 140 | Recovery | = 110.200% |        |          |
| 41) Toluene-d8                | 7.317     | 98             | 231339   | 5.492      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 70 - 130 | Recovery | = 109.800% |        |          |
| 43) trans-1,3-Dichloroprop... | 7.628     | 79             | 23720    | 5.228      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 55 - 130 | Recovery | = 104.600% |        |          |
| 46) 2-Hexanone-d5             | 8.095     | 63             | 106050   | 47.907     | ug/L   | 0.00     |
| Spiked Amount                 | 50.000    | Range 45 - 130 | Recovery | = 95.820%  |        |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217    | 84             | 46761    | 4.760      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 65 - 120 | Recovery | = 95.200%  |        |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.628    | 152            | 90405    | 5.773      | ug/L   | 0.00     |
| Spiked Amount                 | 5.000     | Range 80 - 120 | Recovery | = 115.400% |        |          |
| Target Compounds              |           |                |          |            |        |          |
| 16) Methylene chloride        | 2.507     | 84             | 6988     | 0.430      | ug/L   | 94       |
| 48) 2-Hexanone                | 8.146     | 43             | 18948    | 3.749      | ug/L # | 82       |
| -----                         |           |                |          |            |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV081121\  
Data File : VV021770.D  
Acq On : 11 Aug 2021 12:58  
Operator : SY/MD  
Sample : VV0811WBL01  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK207

Quant Time: Aug 12 03:13:25 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081121WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Aug 12 03:04:55 2021  
Response via : Initial Calibration

