

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081123\
 Data File : VV031792.D
 Acq On : 11 Aug 2023 09:43
 Operator : SY/MD
 Sample : VV0811WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK352

Quant Time: Aug 12 05:19:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 12 05:18:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	183239	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	201428	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	85911	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	67515	4.160	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.200%
7) Chloroethane-d5	1.532	69	67969	4.621	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	2.060	65	38746	4.652	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.000%
20) 2-Butanone-d5	3.825	46	166994	40.281	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.560%
24) Chloroform-d	4.253	84	175831	5.086	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	4.947	65	92440	5.235	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	4.963	84	288577	5.018	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
36) 1,2-Dichloropropane-d6	5.995	67	105208	4.937	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.800%
41) Toluene-d8	7.246	98	219203	4.151	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	7.561	79	27747	4.122	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.400%
46) 2-Hexanone-d5	8.034	63	125088	40.012	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.020%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	69560	4.448	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.000%
66) 1,2-Dichlorobenzene-d4	11.567	152	88484	5.914	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081123\
Data File : VV031792.D
Acq On : 11 Aug 2023 09:43
Operator : SY/MD
Sample : VV0811WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK352

Quant Time: Aug 12 05:19:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 12 05:18:45 2023
Response via : Initial Calibration

