

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV081125\
 Data File : VV038937.D
 Acq On : 11 Aug 2025 10:42
 Operator : SY/MD
 Sample : VV0811WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0811WBL01

Quant Time: Aug 12 01:44:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V080825W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 12 01:42:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.596	168	441418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	774831	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	727563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	330000	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	360350	56.765	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	113.520%
35) Dibromofluoromethane	4.497	113	291655	49.616	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	99.240%
50) Toluene-d8	7.236	98	915451	44.967	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	89.940%
62) 4-Bromofluorobenzene	10.001	95	322195	43.944	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV081125\
Data File : VV038937.D
Acq On : 11 Aug 2025 10:42
Operator : SY/MD
Sample : VV0811WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0811WBL01

Quant Time: Aug 12 01:44:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V080825W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 12 01:42:16 2025
Response via : Initial Calibration

