

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081220\
 Data File : VV017929.D
 Acq On : 12 Aug 2020 15:37
 Operator : SY/MD
 Sample : L3644-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F2L65

Quant Time: Aug 13 05:17:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 13 04:34:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	254338	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	247293	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	108625	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	92885	47.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.72%
7) Chloroethane-d5	1.57	69	74136	46.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.40%
11) 1,1-Dichloroethene-d2	2.12	63	131793	35.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.00%
21) 2-Butanone-d5	3.91	46	131712	103.63	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.63%
24) Chloroform-d	4.38	84	168476	47.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.38%
26) 1,2-Dichloroethane-d4	5.06	65	116007	50.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.12%
32) Benzene-d6	5.07	84	338012	49.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.09	67	110491	49.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.18%
41) Toluene-d8	7.34	98	294082	47.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.96%
43) trans-1,3-Dichloropropene-	7.64	79	47702	47.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.40%
47) 2-Hexanone-d5	8.11	63	72185	100.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.96%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	141264	48.69	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.38%
64) 1,2-Dichlorobenzene-d4	11.65	152	114062	53.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.70%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	507325	224.485	ug/L 99
8) Chloroethane	1.59	64	1320	0.915	ug/L 100
17) trans-1,2-Dichloroethene	2.78	96	22071	12.585	ug/L 95
20) cis-1,2-Dichloroethene	3.94	96	1135179	589.024	ug/L 99
34) Trichloroethene	5.94	95	2626313	1298.650	ug/L 99
46) Tetrachloroethene	8.00	164	428129	288.658	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081220\
Data File : VV017929.D
Acq On : 12 Aug 2020 15:37
Operator : SY/MD
Sample : L3644-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F2L65

Quant Time: Aug 13 05:17:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 13 04:34:59 2020
Response via : Initial Calibration

