

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081222\
 Data File : VW027295.D
 Acq On : 11 Aug 2022 22:57
 Operator : SY/MD
 Sample : N4161-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGC13

Quant Time: Aug 12 05:46:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 02:11:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	164539	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	171698	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69352	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	74494	5.211	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.200%
7) Chloroethane-d5	1.561	69	54474	5.247	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.000%
11) 1,1-Dichloroethene-d2	2.101	63	98142	3.786	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.800%
20) 2-Butanone-d5	3.889	46	88333	36.487	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.980%
24) Chloroform-d	4.342	84	129476	4.939	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
26) 1,2-Dichloroethane-d4	5.031	65	62698	5.020	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.047	84	247270	4.882	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	6.066	67	79465	5.074	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.313	98	206806	4.460	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
43) trans-1,3-Dichloroprop...	7.622	79	24954	4.600	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.088	63	93432	44.506	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.020%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	52359	4.621	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	66421	5.549	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081222\
Data File : VV027295.D
Acq On : 11 Aug 2022 22:57
Operator : SY/MD
Sample : N4161-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGC13

Quant Time: Aug 12 05:46:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 02:11:44 2022
Response via : Initial Calibration

