

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081222\
 Data File : VW027299.D
 Acq On : 12 Aug 2022 00:37
 Operator : SY/MD
 Sample : N4162-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGBW1

Quant Time: Aug 12 05:47:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 02:11:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	158931	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	167531	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	67462	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	73269	5.306	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.200%
7) Chloroethane-d5	1.564	69	53075	5.293	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.800%
11) 1,1-Dichloroethene-d2	2.101	63	97303	3.886	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.800%
20) 2-Butanone-d5	3.889	46	82399	35.236	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.480%
24) Chloroform-d	4.346	84	125731	4.965	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
26) 1,2-Dichloroethane-d4	5.030	65	61389	5.088	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.800%
32) Benzene-d6	5.047	84	239509	4.846	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	6.066	67	75930	4.969	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.400%
41) Toluene-d8	7.313	98	198797	4.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.800%
43) trans-1,3-Dichloroprop...	7.622	79	22863	4.320	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%
46) 2-Hexanone-d5	8.088	63	83261	40.648	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.300%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	49793	4.504	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	62593	5.376	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081222\
Data File : VV027299.D
Acq On : 12 Aug 2022 00:37
Operator : SY/MD
Sample : N4162-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGBW1

Quant Time: Aug 12 05:47:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 02:11:44 2022
Response via : Initial Calibration

