

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081222\
 Data File : VW027306.D
 Acq On : 12 Aug 2022 03:32
 Operator : SY/MD
 Sample : N4162-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGBY2

Quant Time: Aug 12 05:48:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 02:11:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	160023	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	169010	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69254	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	73340	5.275	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.600%
7) Chloroethane-d5	1.564	69	50088	4.961	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.200%
11) 1,1-Dichloroethene-d2	2.101	63	94281	3.739	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.800%
20) 2-Butanone-d5	3.886	46	122332	51.956	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.920%
24) Chloroform-d	4.346	84	121728	4.774	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	5.031	65	60641	4.992	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.800%
32) Benzene-d6	5.047	84	232722	4.668	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
36) 1,2-Dichloropropane-d6	6.066	67	74805	4.853	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.000%
41) Toluene-d8	7.313	98	193930	4.249	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	7.625	79	21432	4.014	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.088	63	84381	40.834	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.660%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	50012	4.484	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	62527	5.231	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081222\
Data File : VV027306.D
Acq On : 12 Aug 2022 03:32
Operator : SY/MD
Sample : N4162-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGBY2

Quant Time: Aug 12 05:48:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 02:11:44 2022
Response via : Initial Calibration

