

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081222\
 Data File : VW027332.D
 Acq On : 12 Aug 2022 15:09
 Operator : SY/MD
 Sample : N4145-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY132

Quant Time: Aug 13 01:38:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 08:00:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	152037	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	168994	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	67645	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	58634	4.439	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.800%
7) Chloroethane-d5	1.561	69	43811	4.567	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.400%
11) 1,1-Dichloroethene-d2	2.101	63	74411	3.106	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.200%
20) 2-Butanone-d5	3.873	46	117714	52.621	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.240%
24) Chloroform-d	4.342	84	110915	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.027	65	52380	4.539	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
32) Benzene-d6	5.047	84	197673	3.965	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
36) 1,2-Dichloropropane-d6	6.066	67	65196	4.230	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.600%
41) Toluene-d8	7.313	98	163578	3.584	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.600%
43) trans-1,3-Dichloroprop...	7.625	79	21428	4.013	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.085	63	80994	39.199	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.400%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	46100	4.134	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	57442	4.920	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081222\
Data File : VV027332.D
Acq On : 12 Aug 2022 15:09
Operator : SY/MD
Sample : N4145-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY132

Quant Time: Aug 13 01:38:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 08:00:46 2022
Response via : Initial Calibration

