

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081222\
 Data File : VV027342.D
 Acq On : 12 Aug 2022 21:40
 Operator : SY/MD
 Sample : N4133-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E11

Quant Time: Aug 13 04:17:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 13 02:46:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	146755	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	155890	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63397	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.298	65	46787	3.669	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.400%
7) Chloroethane-d5	1.561	69	35182	3.799	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.000%
11) 1,1-Dichloroethene-d2	2.098	63	61894	2.677	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.600%#
20) 2-Butanone-d5	3.886	46	123072	56.996	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.000%
24) Chloroform-d	4.342	84	100877	4.314	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.027	65	49124	4.410	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
32) Benzene-d6	5.043	84	169094	3.677	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.600%
36) 1,2-Dichloropropane-d6	6.066	67	57796	4.065	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.200%
41) Toluene-d8	7.313	98	137407	3.264	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.200%#
43) trans-1,3-Dichloroprop...	7.622	79	18661	3.789	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.800%
46) 2-Hexanone-d5	8.088	63	80679	42.329	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.660%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	44696	4.345	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	49734	4.545	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081222\
Data File : VV027342.D
Acq On : 12 Aug 2022 21:40
Operator : SY/MD
Sample : N4133-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5E11

Quant Time: Aug 13 04:17:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 13 02:46:03 2022
Response via : Initial Calibration

