

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW081222\  
 Data File : VW027327.D  
 Acq On : 12 Aug 2022 13:04  
 Operator : SY/MD  
 Sample : N4161-11  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 BGBZ4

Quant Time: Aug 13 01:37:13 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081122WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Aug 12 08:00:46 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc       | Units | Dev(Min) |
|-------------------------------|--------|----------------|----------|------------|-------|----------|
| -----                         |        |                |          |            |       |          |
| Internal Standards            |        |                |          |            |       |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114            | 150830   | 5.000      | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117            | 163156   | 5.000      | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152            | 64252    | 5.000      | ug/L  | 0.00     |
| System Monitoring Compounds   |        |                |          |            |       |          |
| 4) Vinyl Chloride-d3          | 1.301  | 65             | 69468    | 5.301      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | = 106.000% |       |          |
| 7) Chloroethane-d5            | 1.564  | 69             | 49201    | 5.170      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | = 103.400% |       |          |
| 11) 1,1-Dichloroethene-d2     | 2.101  | 63             | 90087    | 3.791      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | = 75.800%  |       |          |
| 20) 2-Butanone-d5             | 3.889  | 46             | 85212    | 38.396     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | = 76.800%  |       |          |
| 24) Chloroform-d              | 4.342  | 84             | 121520   | 5.057      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | = 101.200% |       |          |
| 26) 1,2-Dichloroethane-d4     | 5.031  | 65             | 59902    | 5.232      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | = 104.600% |       |          |
| 32) Benzene-d6                | 5.047  | 84             | 226293   | 4.702      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | = 94.000%  |       |          |
| 36) 1,2-Dichloropropane-d6    | 6.066  | 67             | 72525    | 4.874      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | = 97.400%  |       |          |
| 41) Toluene-d8                | 7.313  | 98             | 188229   | 4.272      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | = 85.400%  |       |          |
| 43) trans-1,3-Dichloroprop... | 7.622  | 79             | 20714    | 4.019      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | = 80.400%  |       |          |
| 46) 2-Hexanone-d5             | 8.088  | 63             | 80117    | 40.162     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | = 80.320%  |       |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84             | 47863    | 4.446      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | = 89.000%  |       |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152            | 62123    | 5.602      | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | = 112.000% |       |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV081222\  
Data File : VV027327.D  
Acq On : 12 Aug 2022 13:04  
Operator : SY/MD  
Sample : N4161-11  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BGBZ4

Quant Time: Aug 13 01:37:13 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR081122WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Aug 12 08:00:46 2022  
Response via : Initial Calibration

