

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
 Data File : VV036892.D
 Acq On : 12 Aug 2024 10:15
 Operator : SY/MD
 Sample : VV0812MBL01
 Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK367

Quant Time: Aug 12 15:04:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	817482	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	759705	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	371411	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	224914	34.447	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	68.900%	
7) Chloroethane-d5	1.536	69	202023	39.235	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	78.480%	
11) 1,1-Dichloroethene-d2	2.063	65	101002	36.163	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	72.320%	
21) 2-Butanone-d5	3.796	46	278324	80.453	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	80.450%	
24) Chloroform-d	4.246	84	508902	43.432	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.860%	
26) 1,2-Dichloroethane-d4	4.937	65	327794	43.511	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.020%	
32) Benzene-d6	4.957	84	990488	43.166	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.340%	
36) 1,2-Dichloropropane-d6	5.986	67	337398	45.578	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	91.160%	
41) Toluene-d8	7.239	98	857773	40.752	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	81.500%	
43) trans-1,3-Dichloroprop...	7.551	79	151762	39.392	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.780%	
47) 2-Hexanone-d5	8.024	63	153057	71.909	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	71.910%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	379352	44.631	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	89.260%	
66) 1,2-Dichlorobenzene-d4	11.554	152	352768	46.611	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.220%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
Data File : VV036892.D
Acq On : 12 Aug 2024 10:15
Operator : SY/MD
Sample : VV0812MBL01
Misc : 5.00G/5 mL/100uL/5.00mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK367

Quant Time: Aug 12 15:04:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 01:11:12 2024
Response via : Initial Calibration

