

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
 Data File : VV036898.D
 Acq On : 12 Aug 2024 12:24
 Operator : SY/MD
 Sample : P3533-01
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AC4

Quant Time: Aug 13 03:22:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 13 02:57:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	754280	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	712484	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	345739	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	269802	44.784	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	89.560%		
7) Chloroethane-d5	1.532	69	212308	44.687	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	89.380%		
11) 1,1-Dichloroethene-d2	2.060	65	113864	44.185	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.360%		
21) 2-Butanone-d5	3.796	46	224279	70.263	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	70.260%		
24) Chloroform-d	4.246	84	468053	43.293	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.580%		
26) 1,2-Dichloroethane-d4	4.937	65	299231	43.048	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.100%		
32) Benzene-d6	4.953	84	966086	44.893	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.780%		
36) 1,2-Dichloropropane-d6	5.985	67	308864	44.488	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	88.980%		
41) Toluene-d8	7.239	98	860462	43.589	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.180%		
43) trans-1,3-Dichloroprop...	7.551	79	137762	38.128	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	76.260%		
47) 2-Hexanone-d5	8.027	63	124912	62.575	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	62.580%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	304401	38.186	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	76.380%		
66) 1,2-Dichlorobenzene-d4	11.554	152	308635	43.808	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.620%		
Target Compounds					Qvalue	
13) Acetone	2.114	43	37276	13.958	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
Data File : VV036898.D
Acq On : 12 Aug 2024 12:24
Operator : SY/MD
Sample : P3533-01
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AC4

Quant Time: Aug 13 03:22:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 13 02:57:00 2024
Response via : Initial Calibration

