

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
 Data File : VV036908.D
 Acq On : 12 Aug 2024 16:24
 Operator : SY/MD
 Sample : P3552-02
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0CJ8

Quant Time: Aug 13 03:25:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 13 02:57:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	1722841	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	1607697	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	785070	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	246411	17.907	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	35.820%#
7) Chloroethane-d5	1.532	69	191466	17.644	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	35.280%#
11) 1,1-Dichloroethene-d2	2.060	65	104634	17.776	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	35.560%#
21) 2-Butanone-d5	3.799	46	235615	32.317	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	32.320%#
24) Chloroform-d	4.243	84	426403	17.268	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	34.540%#
26) 1,2-Dichloroethane-d4	4.937	65	273801	17.245	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	34.500%#
32) Benzene-d6	4.953	84	876374	18.048	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	36.100%#
36) 1,2-Dichloropropane-d6	5.986	67	282009	18.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	36.000%#
41) Toluene-d8	7.240	98	772868	17.351	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	34.700%#
43) trans-1,3-Dichloroprop...	7.551	79	118844	14.577	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	29.160%#
47) 2-Hexanone-d5	8.027	63	128020	28.422	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	28.420%#
56) 1,1,2,2-Tetrachloroeth...	10.149	84	291214	16.190	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	32.380%#
66) 1,2-Dichlorobenzene-d4	11.554	152	282129	17.636	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	35.280%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
Data File : VV036908.D
Acq On : 12 Aug 2024 16:24
Operator : SY/MD
Sample : P3552-02
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0CJ8

Quant Time: Aug 13 03:25:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 13 02:57:00 2024
Response via : Initial Calibration

