

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081224\
 Data File : VW036901.D
 Acq On : 12 Aug 2024 13:29
 Operator : SY/MD
 Sample : P3533-11
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AG6

Quant Time: Aug 13 03:23:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 13 02:57:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	697577	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	664416	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	313193	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	266910	47.905	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.820%
7) Chloroethane-d5	1.532	69	213035	48.485	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.980%
11) 1,1-Dichloroethene-d2	2.060	65	114098	47.874	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.740%
21) 2-Butanone-d5	3.796	46	234664	79.492	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.490%
24) Chloroform-d	4.246	84	459245	45.931	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.860%
26) 1,2-Dichloroethane-d4	4.937	65	300660	46.769	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.540%
32) Benzene-d6	4.953	84	970996	48.386	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.780%
36) 1,2-Dichloropropane-d6	5.985	67	311168	48.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.120%
41) Toluene-d8	7.239	98	864940	46.986	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.980%
43) trans-1,3-Dichloroprop...	7.551	79	139147	41.298	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.600%
47) 2-Hexanone-d5	8.027	63	128127	68.830	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	68.830%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	306902	41.285	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.580%
66) 1,2-Dichlorobenzene-d4	11.554	152	308887	48.399	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
Data File : VV036901.D
Acq On : 12 Aug 2024 13:29
Operator : SY/MD
Sample : P3533-11
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AG6

Quant Time: Aug 13 03:23:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 13 02:57:00 2024
Response via : Initial Calibration

