

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081224\
 Data File : VW036907.D
 Acq On : 12 Aug 2024 16:02
 Operator : SY/MD
 Sample : VIBLK332
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK332

Quant Time: Aug 13 03:24:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 13 02:57:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	703738	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	659276	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	322846	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	196074	34.883	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	69.760%		
7) Chloroethane-d5	1.532	69	174349	39.333	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	78.660%		
11) 1,1-Dichloroethene-d2	2.056	65	89621	37.275	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.540%		
21) 2-Butanone-d5	3.793	46	268037	90.002	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	90.000%		
24) Chloroform-d	4.240	84	434354	43.062	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.120%		
26) 1,2-Dichloroethane-d4	4.934	65	289283	44.606	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.220%		
32) Benzene-d6	4.950	84	834506	41.909	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.820%		
36) 1,2-Dichloropropane-d6	5.982	67	286403	44.583	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.160%		
41) Toluene-d8	7.236	98	717893	39.302	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	78.600%#		
43) trans-1,3-Dichloroprop...	7.548	79	122302	36.581	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	73.160%		
47) 2-Hexanone-d5	8.024	63	151636	82.094	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	82.090%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	347240	47.076	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	94.160%		
66) 1,2-Dichlorobenzene-d4	11.554	152	299711	45.557	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.120%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081224\
Data File : VV036907.D
Acq On : 12 Aug 2024 16:02
Operator : SY/MD
Sample : VIBLK332
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK332

Quant Time: Aug 13 03:24:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080224WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 13 02:57:00 2024
Response via : Initial Calibration

