

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081318\
 Data File : VV007004.D
 Acq On : 13 Aug 2018 14:15
 Operator : SY/MD
 Sample : J4451-23DL 4X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 14 05:36:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 14 05:35:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	223716	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	217264	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	87868	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	56718	3.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.60%
7) Chloroethane-d5	1.58	69	52953	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.40%
11) 1,1-Dichloroethene-d2	2.13	63	82025	3.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.60%
20) 2-Butanone-d5	3.98	46	223137	53.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.46%
24) Chloroform-d	4.41	84	136380	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.09	65	74101	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.20%
32) Benzene-d6	5.10	84	257901	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	6.12	67	91983	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.37	98	194015	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.80%
43) trans-1,3-Dichloropropene-	7.67	79	29440	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
46) 2-Hexanone-d5	8.15	63	107406	39.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.14%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	70430	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	72160	4.72	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.40%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	1.54	94	2458	0.523	ug/L	81
12) 1,1-Dichloroethene	2.14	96	8278	0.515	ug/L	# 1
13) Acetone	2.21	43	37515	17.543	ug/L	76
14) Carbon disulfide	2.32	76	88268	1.664	ug/L	100
15) Methyl Acetate	2.47	43	55626	9.243	ug/L	97
16) Methylene chloride	2.54	84	27571	1.423	ug/L	98
17) Methyl tert-butyl Ether	2.82	73	103750	2.693	ug/L	99
21) 2-Butanone	4.07	43	41258	8.886	ug/L	77
22) cis-1,2-Dichloroethene	3.97	96	29854	1.529	ug/L	100
27) 1,2-Dichloroethane	5.19	62	27388	1.292	ug/L	99
30) Cyclohexane	4.73	56	71932	2.386	ug/L	98
33) Benzene	5.15	78	96279	1.279	ug/L	100
34) Trichloroethene	5.96	95	15464	0.857	ug/L	93
35) Methylcyclohexane	6.18	83	131870	4.551	ug/L	98
40) 4-Methyl-2-pentanone	7.29	43	69547	6.249	ug/L	98
42) Toluene	7.44	91	278180	3.846	ug/L	99
47) Tetrachloroethene	8.03	164	17402	1.189	ug/L	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081318\
Data File : VV007004.D
Acq On : 13 Aug 2018 14:15
Operator : SY/MD
Sample : J4451-23DL 4X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 14 05:36:31 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 14 05:35:33 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
48) 2-Hexanone	8.20	43	61089	7.783	ug/L	99
49) Dibromochloromethane	8.02	129	14726	1.047	ug/L #	11
50) 1,2-Dibromoethane	8.40	107	5731	0.506	ug/L #	97
52) Ethylbenzene	9.06	91	161262	2.233	ug/L	99
53) m,p-xylene	9.19	106	51046	1.896	ug/L	93
62) 1,3-Dichlorobenzene	11.23	146	45880	1.614	ug/L	97
63) 1,4-Dichlorobenzene	11.23	146	45880	1.574	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081318\
Data File : VV007004.D
Acq On : 13 Aug 2018 14:15
Operator : SY/MD
Sample : J4451-23DL 4X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 14 05:36:31 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 14 05:35:33 2018
Response via : Initial Calibration





#6

Bromomethane

Concen: 0.523 ug/L

RT: 1.54 min Scan# 201

Delta R.T. -0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15

Instrument :

MSVOA_V


ClientSampleId :

Tot Ion: 94 Resp: 2458

Ion Ratio Lower Upper




94 100

96 111.0 65.0 120.6



Abundance Ion 94.00 (93.70 to 94.70): VV006999.D (-190) (-)

Ion 96.00 (95.70 to 96.70): VV006999.D (-190) (-)



#12

1,1-Dichloroethene

Concen: 0.515 ug/L

RT: 2.14 min Scan# 390

Delta R.T. -0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15


Tgt Ion: 96 Resp: 8278

Ion Ratio Lower Upper

96 100

61 196.0 111.6 207.4



63 489.8 84.9 157.7#




Abundance Ion 96.00 (95.70 to 96.70): VV006999.D (-376) (-)

Ion 61.05 (60.75 to 61.75): VV006999.D (-376) (-)

Ion 63.00 (62.70 to 63.70): VV006999.D (-376) (-)





#13

Acetone

Concen: 17.543 ug/L

RT: 2.21 min Scan# 411

Delta R.T. 0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15





Instrument :
MSVOA_V
ClientSampleId :

Tot Ion: 43 Resp: 37515

Ion Ratio Lower Upper

43 100

58 21.7 0.0 71.4



#14

Carbon disulfide

Concen: 1.664 ug/L

RT: 2.32 min Scan# 445

Delta R.T. -0.00 min

Lab File: VV007004.D




Acq: 13 Aug 2018 14:15


Tgt Ion: 76 Resp: 88268

Ion Ratio Lower Upper

76 100

78 9.4 7.6 11.4





#15

Methvl Acetate

Concen: 9.243 ug/L

RT: 2.47 min Scan# 492

Delta R.T. 0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15





Instrument :
MSVOA_V
ClientSampleId :

Tot Ion: 43 Resp: 55626

Ion Ratio Lower Upper

43 100

74 29.1 24.6 37.0



#16

Methylene chloride

Concen: 1.423 ug/L

RT: 2.54 min Scan# 513

Delta R.T. 0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15




Tgt Ion: 84 Resp: 27571


Ion Ratio Lower Upper

84 100

86 63.1 45.7 84.9

49 114.1 78.8 146.4









#17
Methvl tert-butvl Ether
Concen: 2.693 ug/L
RT: 2.82 min Scan# 599
Delta R.T. 0.00 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15




Instrument :
MSVOA_V
ClientSampled :


Tot Ion: 73 Resp: 103750
Ion Ratio Lower Upper
73 100
43 17.6 14.4 21.6
57 21.6 17.9 26.9



#21
2-Butanone
Concen: 8.886 ug/L
RT: 4.07 min Scan# 988
Delta R.T. 0.02 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15

Tgt Ion: 43 Resp: 41258
Ion Ratio Lower Upper
43 100
72 13.4 12.4 37.2









#22

cis-1,2-Dichloroethene
 Concen: 1.529 ug/L
 RT: 3.97 min Scan# 957
 Delta R.T. 0.00 min
 Lab File: VV007004.D
 Acq: 13 Aug 2018 14:15

Instrument :
 MSVOA_V
 ClientSampleId :




Tot Ion	96	Resp	29854
Ion	Ratio	Lower	Upper
96	100		
61	130.0	91.1	169.1
68	0.0	0.0	0.0




#27

1,2-Dichloroethane
 Concen: 1.292 ug/L
 RT: 5.19 min Scan# 1337
 Delta R.T. 0.00 min
 Lab File: VV007004.D
 Acq: 13 Aug 2018 14:15

Tgt Ion	62	Resp	27388
Ion	Ratio	Lower	Upper
62	100		
98	9.0	7.6	11.4









#30
Cyclohexane
Concen: 2.386 ug/L
RT: 4.73 min Scan# 1193
Delta R.T. 0.00 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15

Instrument :
MSVOA_V
ClientSampleId :

Tot Ion: 56 Resp: 71932
Ion Ratio Lower Upper
56 100
69 30.3 23.8 35.6
84 87.3 67.9 101.9




Abundance Ion 56.10 (55.80 to 56.80): VV0
Ion 69.15 (68.85 to 69.85): VV0
Ion 84.15 (83.85 to 84.85): VV0






#33
Benzene
Concen: 1.279 ug/L
RT: 5.15 min Scan# 1325
Delta R.T. -0.00 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15

Tgt Ion: 78 Resp: 96279



Abundance Ion 78.10 (77.80 to 78.80): VV0









#34
 Trichloroethene
 Concen: 0.857 ug/L
 RT: 5.96 min Scan# 1578
 Delta R.T. 0.00 min
 Lab File: VV007004.D
 Acq: 13 Aug 2018 14:15

Instrument :
 MSVOA_V
 ClientSampled :

Tot Ion:	95	Resp:	15464
Ion	Ratio	Lower	Upper
95	100		
97	64.9	44.8	83.2
132	100.3	65.5	121.6
130	106.3	66.4	123.4




Abundance Ion 95.00 (94.70 to 95.70): VV006999.D (-1562) (-)
 Ion 96.95 (96.65 to 97.65): VV006999.D (-1562) (-)
 Ion 131.95 (131.65 to 132.65): VV006999.D (-1562) (-)
 Ion 129.95 (129.65 to 130.65): VV006999.D (-1562) (-)






#35
 Methylcyclohexane
 Concen: 4.551 ug/L
 RT: 6.18 min Scan# 1645
 Delta R.T. -0.00 min
 Lab File: VV007004.D
 Acq: 13 Aug 2018 14:15

Tgt Ion:	83	Resp:	131870
Ion	Ratio	Lower	Upper
83	100		
55	86.5	68.1	102.1
98	49.8	37.9	56.9



Abundance Ion 83.15 (82.85 to 83.85): VV006999.D (-1633) (-)
 Ion 55.10 (54.80 to 55.80): VV006999.D (-1633) (-)
 Ion 98.15 (97.85 to 98.85): VV006999.D (-1633) (-)









#40
4-Methyl-2-pentanone
Concen: 6.249 ug/L
RT: 7.29 min Scan# 1991
Delta R.T. 0.01 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15

Instrument :
MSVOA_V
ClientSampled :

Tot Ion: 43 Resp: 69547
Ion Ratio Lower Upper
43 100
58 36.9 30.7 46.1
100 14.1 11.2 16.8




Abundance Ion 43.05 (42.75 to 43.75): VV0
Ion 58.05 (57.75 to 58.75): VV0
Ion 100.15 (99.85 to 100.85): VV0






#42
Toluene
Concen: 3.846 ug/L
RT: 7.44 min Scan# 2036
Delta R.T. 0.00 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15

Tgt Ion: 91 Resp: 278180
Ion Ratio Lower Upper
91 100
92 57.8 40.2 74.6



Abundance Ion 91.15 (90.85 to 91.85): VV0
Ion 92.15 (91.85 to 92.85): VV0









#47
Tetrachloroethene
Concen: 1.189 ug/L
RT: 8.03 min Scan# 2219
Delta R.T. 0.00 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15




Instrument :
MSVOA_V
ClientSampleId :


Tot Ion:164 Resp: 17402
Ion Ratio Lower Upper
164 100
129 95.7 65.9 122.3
131 88.5 64.2 119.2
166 130.4 89.4 166.0



#48
2-Hexanone
Concen: 7.783 ug/L
RT: 8.20 min Scan# 2273
Delta R.T. 0.00 min
Lab File: VV007004.D
Acq: 13 Aug 2018 14:15

Tgt Ion: 43 Resp: 61089
Ion Ratio Lower Upper
43 100
58 55.5 44.2 66.2
57 20.8 14.8 22.2
100 11.0 8.8 13.2





#49

Dibromochloromethane

Concen: 1.047 ug/L

RT: 8.02 min Scan# 2218

Delta R.T. -0.27 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15

Instrument :

MSVOA_V


ClientSampleId :

Tot Ion:129 Resp: 14726

Ion Ratio Lower Upper

129 100

127 0.0 53.3 98.9#



Abundance Ion 128.95 (128.65 to 129.65): V

Ion 126.90 (126.60 to 127.60): V

8.02

10000

8000



6000

4000

2000

0

Time-->



#50

1,2-Dibromoethane

Concen: 0.506 ug/L

RT: 8.40 min Scan# 2336

Delta R.T. -0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15


Tgt Ion:107 Resp: 5731

Ion Ratio Lower Upper

107 100

109 88.7 64.1 119.1

188 3.8 4.0 6.0#



Abundance Ion 106.95 (106.65 to 107.65): V

Ion 109.00 (108.70 to 109.70): V

Ion 187.90 (187.60 to 188.60): V

8.40

4000


3000


2000

1000

0

Time-->





#52

Ethylbenzene

Concen: 2.233 ug/L

RT: 9.06 min Scan# 2541

Delta R.T. -0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15

Instrument :

MSVOA_V


ClientSampleId :

Tot Ion: 91 Resp: 161262




Ion Ratio Lower Upper

91 100

106 29.7 21.3 39.5



Abundance Ion 91.15 (90.85 to 91.85): VV006999.D (-2527) (-)



#53

m,p-xylene

Concen: 1.896 ug/L

RT: 9.19 min Scan# 2582

Delta R.T. 0.01 min

Lab File: VV007004.D


Acq: 13 Aug 2018 14:15

Tgt Ion: 106 Resp: 51046



Ion Ratio Lower Upper


106 100

91 192.3 141.6 263.0



Abundance Ion 106.15 (105.85 to 106.85): VV006999.D (-2567) (-)





#62

1,3-Dichlorobenzene

Concen: 1.614 ug/L

RT: 11.23 min Scan# 3217

Delta R.T. 0.00 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15

Instrument :

MSVOA_V

ClientSampleId :


Tot Ion:146 Resp: 45880

Ion Ratio Lower Upper




146 100

111 36.4 28.1 52.3

148 64.3 44.2 82.2



Abundance Ion 145.95 (145.65 to 146.65): V
Ion 111.00 (110.70 to 111.70): V
Ion 148.00 (147.70 to 148.70): V



#63

1,4-Dichlorobenzene

Concen: 1.574 ug/L

RT: 11.23 min Scan# 3217

Delta R.T. -0.09 min

Lab File: VV007004.D

Acq: 13 Aug 2018 14:15


Tgt Ion:146 Resp: 45880

Ion Ratio Lower Upper

146 100

111 36.4 26.6 49.4

148 64.3 46.4 86.2



Abundance Ion 145.95 (145.65 to 146.65): V
Ion 111.00 (110.70 to 111.70): V
Ion 148.00 (147.70 to 148.70): V

