

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV081325\
 Data File : VV038952.D
 Acq On : 13 Aug 2025 15:02
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Aug 14 02:45:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V081325W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 14 02:40:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	434539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	824447	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	729910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	302852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	355395	59.300	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 118.600%#	
35) Dibromofluoromethane	4.500	113	309373	53.267	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 106.540%	
50) Toluene-d8	7.236	98	1006159	49.728	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 99.460%	
62) 4-Bromofluorobenzene	10.001	95	308216	41.925	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 83.860%#	
Target Compounds						Qvalue
20) Methylene Chloride	2.458	84	9899	1.158	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV081325\
Data File : VV038952.D
Acq On : 13 Aug 2025 15:02
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Aug 14 02:45:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V081325W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 14 02:40:30 2025
Response via : Initial Calibration

