

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081419\
 Data File : VV012245.D
 Acq On : 14 Aug 2019 19:59
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00540

Quant Time: Aug 15 23:33:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 15 23:30:01 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	5.000	4.847	3.1	85	0.00
3 T	Chloromethane	5.000	4.773	4.5	87	0.00
4 S	Vinyl Chloride-d3	5.000	4.090	18.2	81	0.00
5 T	Vinyl chloride	5.000	4.810	3.8	86	0.00
6 T	Bromomethane	5.000	3.259	34.8#	60	0.00
7 S	Chloroethane-d5	5.000	4.542	9.2	87	0.00
8 T	Chloroethane	5.000	4.834	3.3	85	0.00
9 T	Trichlorofluoromethane	5.000	4.690	6.2	85	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	4.893	2.1	92	0.00
11 S	1,1-Dichloroethene-d2	5.000	4.468	10.6	90	0.00
12 T	1,1-Dichloroethene	5.000	4.964	0.7	97	0.00
13 T	Acetone	50.000	46.117	7.8	90	0.00
14 T	Carbon disulfide	5.000	3.818	23.6	74	0.00
15 T	Methyl Acetate	5.000	3.908	21.8	79	0.00
16 T	Methylene chloride	5.000	3.693	26.1	74	0.00
17 T	Methyl tert-butyl Ether	5.000	4.533	9.3	85	0.00
18 T	trans-1,2-Dichloroethene	5.000	4.204	15.9	78	0.00
19 T	1,1-Dichloroethane	5.000	5.151	-3.0	94	0.00
20 S	2-Butanone-d5	50.000	55.069	-10.1	101	0.00
21 T	2-Butanone	50.000	56.462	-12.9	97	0.00
22 T	cis-1,2-Dichloroethene	5.000	4.811	3.8	85	0.00
23 T	Bromochloromethane	5.000	5.190	-3.8	92	0.00
24 S	Chloroform-d	5.000	4.719	5.6	90	0.00
25 T	Chloroform	5.000	4.804	3.9	92	0.00
26 S	1,2-Dichloroethane-d4	5.000	4.759	4.8	89	0.00
27 T	1,2-Dichloroethane	5.000	5.124	-2.5	91	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	95	0.00
29 T	1,1,1-Trichloroethane	5.000	4.909	1.8	89	0.00
30 T	Cyclohexane	5.000	4.681	6.4	84	0.00
31 T	Carbon tetrachloride	5.000	4.879	2.4	88	0.00
32 S	Benzene-d6	5.000	4.428	11.4	85	0.00
33 T	Benzene	5.000	4.998	0.0	89	0.00
34 T	Trichloroethene	5.000	4.690	6.2	87	0.00
35 T	Methylcyclohexane	5.000	4.654	6.9	83	0.00
36 S	1,2-Dichloropropane-d6	5.000	4.542	9.2	89	0.00
37 T	1,2-Dichloropropane	5.000	4.901	2.0	88	0.00
38 T	Bromodichloromethane	5.000	5.002	-0.0	92	0.00
39 T	cis-1,3-Dichloropropene	5.000	4.910	1.8	86	0.00
40 T	4-Methyl-2-pentanone	50.000	57.983	-16.0	100	0.00
41 S	Toluene-d8	5.000	4.553	8.9	86	0.00
42 T	Toluene	5.000	5.031	-0.6	88	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	4.585	8.3	89	0.00
44 T	trans-1,3-Dichloropropene	5.000	5.094	-1.9	93	0.00
45 T	1,1,2-Trichloroethane	5.000	5.312	-6.2	95	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081419\
 Data File : VV012245.D
 Acq On : 14 Aug 2019 19:59
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00540

Quant Time: Aug 15 23:33:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 15 23:30:01 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	50.000	55.553	-11.1	98	0.00
47 T	Tetrachloroethene	5.000	4.831	3.4	87	0.00
48 T	2-Hexanone	50.000	59.270	-18.5	100	0.00
49 T	Dibromochloromethane	5.000	5.113	-2.3	92	0.00
50 T	1,2-Dibromoethane	5.000	5.157	-3.1	93	0.00
51 T	Chlorobenzene	5.000	4.894	2.1	88	0.00
52 T	Ethylbenzene	5.000	4.924	1.5	87	0.00
53 T	m,p-xylene	5.000	4.998	0.0	87	0.00
54 T	o-xylene	5.000	4.967	0.7	87	0.00
55 T	Styrene	5.000	5.196	-3.9	88	0.00
56 T	Isopropylbenzene	5.000	4.938	1.2	87	0.00
57 S	1,1,2,2-Tetrachloroethane-d	5.000	5.060	-1.2	96	0.00
58 T	1,1,2,2-Tetrachloroethane	5.000	5.334	-6.7	95	0.00
59	1,2,3-Trichloropropane	5.000	5.495	-9.9	98	0.00
60 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	0.00
61 T	Bromoform	5.000	4.988	0.2	91	0.00
62 T	1,3-Dichlorobenzene	5.000	4.730	5.4	87	0.00
63 T	1,4-Dichlorobenzene	5.000	4.684	6.3	87	0.00
64 S	1,2-Dichlorobenzene-d4	5.000	4.403	11.9	89	0.00
65 T	1,2-Dichlorobenzene	5.000	4.876	2.5	90	0.00
66 T	1,2-Dibromo-3-chloropropane	5.000	4.940	1.2	100	0.00
67	1,3,5-Trichlorobenzene	5.000	4.538	9.2	85	0.00
68 T	1,2,4-trichlorobenzene	5.000	4.661	6.8	86	0.00
69	Naphthalene	5.000	5.079	-1.6	92	0.00
70 T	1,2,3-Trichlorobenzene	5.000	4.982	0.4	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0