

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081519\
 Data File : VV012259.D
 Acq On : 15 Aug 2019 20:22
 Operator : SY/MD
 Sample : K4373-06 50X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF4

Quant Time: Aug 16 05:14:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 16 04:21:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	182993	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	173780	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	78881	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40268	3.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.40%
7) Chloroethane-d5	1.58	69	31779	3.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.80%
11) 1,1-Dichloroethene-d2	2.13	63	63025	3.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.20%
20) 2-Butanone-d5	3.96	46	124709	51.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.44%
24) Chloroform-d	4.40	84	105363	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.08	65	52734	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.10	84	207260	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.12	67	60004	4.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.60%
41) Toluene-d8	7.36	98	180503	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	7.66	79	22347	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	8.14	63	78858	49.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.14%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	44385	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	70464	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081519\
Data File : VV012259.D
Acq On : 15 Aug 2019 20:22
Operator : SY/MD
Sample : K4373-06 50X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF4

Quant Time: Aug 16 05:14:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR081319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 16 04:21:37 2019
Response via : Initial Calibration

