

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081522\
 Data File : VV027392.D
 Acq On : 15 Aug 2022 16:47
 Operator : SY/MD
 Sample : N4192-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB6

Quant Time: Aug 16 02:38:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 16 00:24:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	177270	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	183865	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	79600	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	49607	3.221	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	64.400%	
7) Chloroethane-d5	1.561	69	40186	3.593	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	71.800%	
11) 1,1-Dichloroethene-d2	2.101	63	63366	2.269	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	45.400%#	
20) 2-Butanone-d5	3.873	46	126300	48.422	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	96.840%	
24) Chloroform-d	4.342	84	109505	3.877	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	77.600%	
26) 1,2-Dichloroethane-d4	5.027	65	54108	4.021	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.400%	
32) Benzene-d6	5.047	84	196790	3.628	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	72.600%	
36) 1,2-Dichloropropane-d6	6.066	67	65699	3.918	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	78.400%	
41) Toluene-d8	7.313	98	158663	3.195	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	64.000%#	
43) trans-1,3-Dichloroprop...	7.622	79	21561	3.712	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.200%	
46) 2-Hexanone-d5	8.085	63	87623	38.977	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.960%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	47981	3.955	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	79.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	61012	4.441	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.800%	
Target Compounds						Qvalue
16) Methylene chloride	2.503	84	7588	0.370	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081522\
Data File : VV027392.D
Acq On : 15 Aug 2022 16:47
Operator : SY/MD
Sample : N4192-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB6

Quant Time: Aug 16 02:38:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 16 00:24:20 2022
Response via : Initial Calibration

