

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081523\
 Data File : VW031811.D
 Acq On : 15 Aug 2023 15:12
 Operator : SY/MD
 Sample : 03968-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A48B1

Quant Time: Aug 16 02:13:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 16 02:07:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	338259	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	376656	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	170614	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	121178	35.141	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	70.280%		
7) Chloroethane-d5	1.532	69	125456	39.700	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	79.400%		
11) 1,1-Dichloroethene-d2	2.060	65	58932	36.921	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	73.840%		
21) 2-Butanone-d5	3.793	46	240043	100.600	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	100.600%		
24) Chloroform-d	4.256	84	324927	45.831	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.660%		
26) 1,2-Dichloroethane-d4	4.947	65	209325	47.453	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.900%		
32) Benzene-d6	4.966	84	493965	41.415	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.840%		
36) 1,2-Dichloropropane-d6	5.995	67	197940	44.180	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	88.360%		
41) Toluene-d8	7.249	98	362578	36.268	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	72.540%#		
43) trans-1,3-Dichloroprop...	7.558	79	73024	37.920	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	75.840%		
47) 2-Hexanone-d5	8.031	63	133977	79.744	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.740%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	257691	41.745	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	83.480%		
66) 1,2-Dichlorobenzene-d4	11.567	152	177223	51.303	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	102.600%		
Target Compounds					Qvalue	
13) Acetone	2.114	43	90821	39.765	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081523\
Data File : VV031811.D
Acq On : 15 Aug 2023 15:12
Operator : SY/MD
Sample : 03968-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A48B1

Quant Time: Aug 16 02:13:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 16 02:07:29 2023
Response via : Initial Calibration

