

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081618\
 Data File : VV007035.D
 Acq On : 16 Aug 2018 11:09
 Operator : SY/MD
 Sample : VV0816WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK61

Quant Time: Aug 17 01:21:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 17 01:13:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	234788	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	220608	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	86355	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	68875	4.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.80%
7) Chloroethane-d5	1.58	69	57420	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.40%
11) 1,1-Dichloroethene-d2	2.13	63	91650	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	3.97	46	144384	33.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	66.26%
24) Chloroform-d	4.40	84	128385	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
26) 1,2-Dichloroethane-d4	5.09	65	64620	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
32) Benzene-d6	5.10	84	254240	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.12	67	84373	4.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.20%
41) Toluene-d8	7.36	98	187855	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloropropene-	7.67	79	25006	3.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.20%
46) 2-Hexanone-d5	8.15	63	69776	25.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	50.62%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	55854	3.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	64049	4.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081618\
Data File : VV007035.D
Acq On : 16 Aug 2018 11:09
Operator : SY/MD
Sample : VV0816WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK61

Quant Time: Aug 17 01:21:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 17 01:13:44 2018
Response via : Initial Calibration

