

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081621\
 Data File : VV021784.D
 Acq On : 16 Aug 2021 11:41
 Operator : SY/MD
 Sample : IBLK200
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 IBLK200

Quant Time: Aug 17 04:15:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 17 04:08:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	451481	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	419604	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	206179	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	169413	47.999	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	96.000%		
7) Chloroethane-d5	1.561	69	126782	47.600	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	95.200%		
11) 1,1-Dichloroethene-d2	2.104	63	206235	36.708	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	73.420%		
21) 2-Butanone-d5	3.870	46	222181	101.215	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	101.210%		
24) Chloroform-d	4.346	84	283159	46.305	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.620%		
26) 1,2-Dichloroethane-d4	5.030	65	174291	49.550	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.100%		
32) Benzene-d6	5.047	84	602554	49.937	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.880%		
36) 1,2-Dichloropropane-d6	6.069	67	187658	49.941	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.880%		
41) Toluene-d8	7.316	98	548575	49.188	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.380%		
43) trans-1,3-Dichloroprop...	7.622	79	89975	48.654	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.300%		
47) 2-Hexanone-d5	8.088	63	116635	94.706	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	94.710%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	212263	45.585	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.160%		
66) 1,2-Dichlorobenzene-d4	11.625	152	211997	49.684	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	99.360%		
Target Compounds						Qvalue
14) Carbon disulfide	2.291	76	8103	0.902	ug/L	97
16) Methylene chloride	2.503	84	7909	2.131	ug/L	96
25) Chloroform	4.378	83	6163	1.027	ug/L	97
48) 2-Hexanone	8.143	43	11611	3.333	ug/L #	86
70) 1,2,4-trichlorobenzene	13.265	180	5357	1.316	ug/L	97
72) 1,2,3-Trichlorobenzene	13.750	180	6485	1.586	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081621\
Data File : VV021784.D
Acq On : 16 Aug 2021 11:41
Operator : SY/MD
Sample : IBLK200
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
IBLK200

Quant Time: Aug 17 04:15:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 17 04:08:59 2021
Response via : Initial Calibration

