

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081622\
 Data File : VV027410.D
 Acq On : 16 Aug 2022 16:24
 Operator : SY/MD
 Sample : N4192-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Aug 17 00:30:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 16 00:24:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	135538	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	142520	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62211	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	39700	3.371	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.400%
7) Chloroethane-d5	1.564	69	34627	4.049	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.000%
11) 1,1-Dichloroethene-d2	2.101	63	64164	3.005	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.000%
20) 2-Butanone-d5	3.892	46	119004	59.673	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.340%
24) Chloroform-d	4.342	84	96357	4.462	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	5.027	65	51659	5.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.043	84	174646	4.154	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.066	67	61102	4.700	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.313	98	135105	3.510	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.200%
43) trans-1,3-Dichloroprop...	7.622	79	20121	4.469	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.088	63	88061	50.536	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.080%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	47079	5.006	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	51778	4.822	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.400%
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	17198	1.097	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081622\
Data File : VV027410.D
Acq On : 16 Aug 2022 16:24
Operator : SY/MD
Sample : N4192-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Aug 17 00:30:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 16 00:24:20 2022
Response via : Initial Calibration

