

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081718\
 Data File : VV007056.D
 Acq On : 17 Aug 2018 15:56
 Operator : SY/MD
 Sample : J4488-03
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0355

Quant Time: Aug 18 04:21:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 18 03:36:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	211478	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	206337	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78860	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	77878	5.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.00%
7) Chloroethane-d5	1.58	69	66946	5.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.60%
11) 1,1-Dichloroethene-d2	2.13	63	149791	5.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	117.00%
20) 2-Butanone-d5	3.97	46	201623	51.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.72%
24) Chloroform-d	4.41	84	157060	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.40%
26) 1,2-Dichloroethane-d4	5.09	65	80161	5.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.60%
32) Benzene-d6	5.10	84	314638	5.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.80%
36) 1,2-Dichloropropane-d6	6.12	67	104066	5.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.80%
41) Toluene-d8	7.37	98	225126	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	7.67	79	30313	4.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.20%
46) 2-Hexanone-d5	8.15	63	81739	31.71	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.42%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	73722	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	80390	5.86	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.20%

Target Compounds

						Ovalue
12) 1,1-Dichloroethene	2.14	96	86655	5.71	ug/L	92
13) Acetone	2.21	43	7561	3.74	ug/L	91
19) 1,1-Dichloroethane	3.23	63	738947	25.17	ug/L	99
25) Chloroform	4.43	83	7255	0.23	ug/L	95
27) 1,2-Dichloroethane	5.19	62	6674	0.33	ug/L	97
29) 1,1,1-Trichloroethane	4.66	97	5505856	200.12	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081718\
Data File : VV007056.D
Acq On : 17 Aug 2018 15:56
Operator : SY/MD
Sample : J4488-03
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0355

Quant Time: Aug 18 04:21:26 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 18 03:36:46 2018
Response via : Initial Calibration

