

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081718\
 Data File : VV007057.D
 Acq On : 17 Aug 2018 16:23
 Operator : SY/MD
 Sample : J4488-04
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0356

Quant Time: Aug 18 04:24:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 18 03:36:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	219860	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	214766	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	83435	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82585	5.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	116.40%
7) Chloroethane-d5	1.58	69	73461	6.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.00%
11) 1,1-Dichloroethene-d2	2.13	63	115413	4.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.60%
20) 2-Butanone-d5	3.97	46	217226	53.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.44%
24) Chloroform-d	4.41	84	165987	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.40%
26) 1,2-Dichloroethane-d4	5.09	65	85201	5.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.20%
32) Benzene-d6	5.10	84	319960	5.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.00%
36) 1,2-Dichloropropane-d6	6.12	67	107076	5.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.60%
41) Toluene-d8	7.37	98	238080	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
43) trans-1,3-Dichloropropene-	7.67	79	31934	4.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.40%
46) 2-Hexanone-d5	8.15	63	94402	35.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.36%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	77301	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	83980	5.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.80%

Target Compounds					Ovalue
13) Acetone	2.20	43	7411	3.53 ug/L	90
29) 1,1,1-Trichloroethane	4.66	97	11610	0.41 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081718\
Data File : VV007057.D
Acq On : 17 Aug 2018 16:23
Operator : SY/MD
Sample : J4488-04
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0356

Quant Time: Aug 18 04:24:21 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 18 03:36:46 2018
Response via : Initial Calibration

