

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081718\
 Data File : VV007046.D
 Acq On : 17 Aug 2018 11:13
 Operator : SY/MD
 Sample : VV0817WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK62

Quant Time: Aug 18 03:42:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 18 03:36:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	224806	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	216839	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	83183	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82377	5.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.60%
7) Chloroethane-d5	1.58	69	72458	5.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.80%
11) 1,1-Dichloroethene-d2	2.13	63	112444	4.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.60%
20) 2-Butanone-d5	3.98	46	226020	54.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.32%
24) Chloroform-d	4.40	84	166952	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.40%
26) 1,2-Dichloroethane-d4	5.09	65	83991	5.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.00%
32) Benzene-d6	5.10	84	327336	5.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.60%
36) 1,2-Dichloropropane-d6	6.12	67	108448	5.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.00%
41) Toluene-d8	7.36	98	252161	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.40%
43) trans-1,3-Dichloropropene-	7.67	79	33393	5.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.80%
46) 2-Hexanone-d5	8.15	63	97079	35.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.66%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	75612	5.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	85417	5.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081718\
Data File : VV007046.D
Acq On : 17 Aug 2018 11:13
Operator : SY/MD
Sample : VV0817WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK62

Quant Time: Aug 18 03:42:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 18 03:36:46 2018
Response via : Initial Calibration

