

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081720\
 Data File : VV018038.D
 Acq On : 17 Aug 2020 20:18
 Operator : SY/MD
 Sample : L3703-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Aug 18 03:38:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 18 01:23:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	261845	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	258971	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	123495	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	86826	43.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.92%
7) Chloroethane-d5	1.58	69	75483	46.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.36%
11) 1,1-Dichloroethene-d2	2.12	63	132106	34.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.14%
21) 2-Butanone-d5	3.91	46	123127	94.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.10%
24) Chloroform-d	4.38	84	162080	44.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.14%
26) 1,2-Dichloroethane-d4	5.06	65	111993	47.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.82%
32) Benzene-d6	5.08	84	324854	45.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.04%
36) 1,2-Dichloropropane-d6	6.09	67	107085	45.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.78%
41) Toluene-d8	7.34	98	284128	44.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.54%
43) trans-1,3-Dichloropropene-	7.64	79	47502	44.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.78%
47) 2-Hexanone-d5	8.11	63	59284	79.17	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.17%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	129773	42.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.42%
64) 1,2-Dichlorobenzene-d4	11.65	152	119263	49.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081720\
Data File : VV018038.D
Acq On : 17 Aug 2020 20:18
Operator : SY/MD
Sample : L3703-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Aug 18 03:38:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 18 01:23:44 2020
Response via : Initial Calibration

