

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081720\
 Data File : VV018015.D
 Acq On : 17 Aug 2020 10:42
 Operator : SY/MD
 Sample : VV0817WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK82

Quant Time: Aug 18 01:38:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 18 01:23:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	244753	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	246075	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	117220	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94273	50.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.96%
7) Chloroethane-d5	1.58	69	80288	52.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.10%
11) 1,1-Dichloroethene-d2	2.12	63	141244	39.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.08%
21) 2-Butanone-d5	3.91	46	120202	98.28	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.28%
24) Chloroform-d	4.38	84	171836	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.10%
26) 1,2-Dichloroethane-d4	5.06	65	117890	53.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.78%
32) Benzene-d6	5.07	84	344278	50.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.54%
36) 1,2-Dichloropropane-d6	6.09	67	112745	50.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.70%
41) Toluene-d8	7.34	98	304797	49.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.96%
43) trans-1,3-Dichloropropene-	7.64	79	51480	51.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.38%
47) 2-Hexanone-d5	8.11	63	56953	80.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.05%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	132551	45.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.82%
64) 1,2-Dichlorobenzene-d4	11.65	152	126499	54.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.66%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081720\
Data File : VV018015.D
Acq On : 17 Aug 2020 10:42
Operator : SY/MD
Sample : VV0817WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK82

Quant Time: Aug 18 01:38:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 18 01:23:44 2020
Response via : Initial Calibration

