


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV081720\
 Data File : VV018012.D
 Acq On : 17 Aug 2020 09:21
 Operator : SY/MD
 Sample : BFB75
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB75

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Title : VOC Analysis
 Last Update : Tue Aug 18 01:23:44 2020



AutoFind: Scans 2800, 2801, 2802; Background Corrected with Scan 2790

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	5253	PASS
75	95	30	80	52.6	14224	PASS
95	95	100	100	100.0	27042	PASS
96	95	5	9	6.7	1822	PASS
173	174	0.00	2	0.6	125	PASS
174	95	50	120	82.4	22293	PASS
175	174	5	9	8.6	1914	PASS
176	174	95	101	95.1	21202	PASS
177	176	5	9	6.3	1342	PASS

m/z	Abundance
35.90	225.0
37.00	1206.0
38.00	1193.0
39.10	525.0
39.90	211.0
44.00	501.0
44.90	277.0
47.00	400.0
47.90	176.0
48.90	1170.0
50.00	5123.0
51.00	1771.0
51.90	107.0
54.90	122.0
56.00	450.0
57.00	705.0
60.00	305.0
61.00	1397.0
62.00	1292.0
63.00	1088.0
64.00	191.0
64.90	357.0
66.90	129.0
68.00	2609.0
69.00	2908.0
69.90	190.0
71.90	156.0
73.00	1248.0
74.00	4378.0
75.00	13806.0
76.00	1255.0
77.00	248.0
77.90	112.0
78.90	1041.0
79.90	307.0
80.90	950.0
81.80	279.0
86.90	1106.0
87.90	1085.0
90.80	122.0
92.00	911.0
92.90	1286.0
94.00	3150.0
95.00	26168.0
96.00	1817.0
103.80	109.0
105.80	108.0
116.00	117.0
116.80	272.0
117.70	132.0
118.80	195.0
127.70	103.0
140.90	258.0
142.90	362.0
144.90	132.0
171.90	414.0
173.00	375.0
173.90	20744.0
174.90	1803.0
175.90	19944.0
176.90	1285.0

Instrument :
MSVOA_V
ClientSampleId :
BFB75

m/z	Abundance
36.00	247.0
37.00	1297.0
38.00	1253.0
39.00	610.0
39.90	205.0
44.00	485.0
44.90	296.0
47.00	410.0
47.90	178.0
49.00	1246.0
50.00	5421.0
51.00	1818.0
54.90	134.0
56.00	449.0
57.00	705.0
59.90	300.0
61.00	1421.0
62.00	1378.0
63.00	1141.0
64.00	199.0
64.90	421.0
66.90	131.0
68.00	2816.0
69.00	3078.0
69.90	209.0
71.90	160.0
73.00	1459.0
74.00	4647.0
75.00	14794.0
76.00	1275.0
77.00	214.0
78.00	122.0
78.90	1031.0
79.90	346.0
80.90	1010.0
81.80	279.0
87.00	1227.0
87.90	1119.0
90.90	119.0
91.90	925.0
93.00	1420.0
94.00	3452.0
95.00	27984.0
96.00	1891.0
103.80	140.0
105.80	102.0
116.80	266.0
117.70	150.0
118.80	231.0
140.90	257.0
142.90	377.0
144.90	143.0
171.90	450.0
173.90	23032.0
174.90	1981.0
175.90	21840.0
176.90	1374.0

Instrument :
MSVOA_V
ClientSampleId :
BFB75

m/z	Abundance
36.00	277.0
37.00	1231.0
38.00	1182.0
39.00	664.0
40.00	211.0
43.90	497.0
45.00	280.0
47.00	366.0
47.90	177.0
49.00	1152.0
50.00	5215.0
51.00	1712.0
54.90	138.0
56.00	428.0
57.00	645.0
59.90	283.0
61.00	1309.0
62.00	1419.0
63.00	1105.0
64.00	197.0
65.00	446.0
67.00	103.0
68.00	2764.0
69.00	2923.0
69.90	208.0
71.90	149.0
73.00	1427.0
74.00	4624.0
75.00	14074.0
76.00	1134.0
76.90	214.0
77.90	120.0
78.90	944.0
79.90	362.0
80.90	962.0
81.80	243.0
87.00	1255.0
87.90	1050.0
90.90	122.0
91.90	792.0
93.00	1377.0
94.00	3280.0
95.00	26976.0
96.00	1759.0
103.80	145.0
116.90	249.0
117.70	146.0
118.80	229.0
140.90	225.0
142.90	342.0
144.90	126.0
171.90	432.0
173.90	23104.0
174.90	1958.0
175.90	21824.0
176.90	1369.0

Instrument :
MSVOA_V
ClientSampleId :
BFB75

Scan 2790 (10.061 min): VV018012.D
BFB75

m/z	Abundance
39.90	156.0
43.90	394.0

Instrument :
MSVOA_V
ClientSampleId :
BFB75