

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081721\
 Data File : VW021841.D
 Acq On : 18 Aug 2021 02:11
 Operator : SY/MD
 Sample : M3422-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKD6

Quant Time: Aug 18 03:39:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 18 03:33:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	415776	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	386370	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	177619	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	111344	34.256	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.520%
7) Chloroethane-d5	1.558	69	98080	39.986	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.980%
11) 1,1-Dichloroethene-d2	2.105	63	129060	24.944	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	49.880%#
21) 2-Butanone-d5	3.873	46	227255	112.417	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.420%
24) Chloroform-d	4.349	84	248433	44.115	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.240%
26) 1,2-Dichloroethane-d4	5.031	65	156719	48.381	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.760%
32) Benzene-d6	5.050	84	492822	44.356	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.720%
36) 1,2-Dichloropropane-d6	6.072	67	167600	48.440	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.880%
41) Toluene-d8	7.317	98	430365	41.908	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.820%
43) trans-1,3-Dichloroprop...	7.625	79	70209	41.231	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.460%
47) 2-Hexanone-d5	8.088	63	116600	102.822	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.820%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	209246	48.802	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	184112	50.087	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081721\
Data File : VV021841.D
Acq On : 18 Aug 2021 02:11
Operator : SY/MD
Sample : M3422-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKD6

Quant Time: Aug 18 03:39:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 18 03:33:12 2021
Response via : Initial Calibration

