

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081721\
 Data File : VW021843.D
 Acq On : 18 Aug 2021 02:59
 Operator : SY/MD
 Sample : M3422-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKD8

Quant Time: Aug 18 03:39:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 18 03:33:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	408192	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	378085	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	177334	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	110964	34.773	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.540%
7) Chloroethane-d5	1.558	69	97760	40.596	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.200%
11) 1,1-Dichloroethene-d2	2.105	63	136513	26.875	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.740%#
21) 2-Butanone-d5	3.873	46	215931	108.800	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.800%
24) Chloroform-d	4.349	84	251164	45.429	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.860%
26) 1,2-Dichloroethane-d4	5.034	65	155405	48.867	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.740%
32) Benzene-d6	5.050	84	492272	45.277	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.560%
36) 1,2-Dichloropropane-d6	6.072	67	165237	48.803	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.600%
41) Toluene-d8	7.317	98	423330	42.126	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.260%
43) trans-1,3-Dichloroprop...	7.625	79	68046	40.837	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.680%
47) 2-Hexanone-d5	8.088	63	118106	106.432	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.430%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	220429	52.537	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.080%
66) 1,2-Dichlorobenzene-d4	11.625	152	184083	50.159	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081721\
Data File : VV021843.D
Acq On : 18 Aug 2021 02:59
Operator : SY/MD
Sample : M3422-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKD8

Quant Time: Aug 18 03:39:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 18 03:33:12 2021
Response via : Initial Calibration

