

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081721\
 Data File : VW021846.D
 Acq On : 18 Aug 2021 04:11
 Operator : SY/MD
 Sample : M3422-22
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 DBKE1

Quant Time: Aug 18 04:38:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 18 03:33:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	408943	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	385321	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	193772	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	107751	33.704	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	67.400%		
7) Chloroethane-d5	1.558	69	93675	38.829	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery =	77.660%		
11) 1,1-Dichloroethene-d2	2.101	63	135559	26.638	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	53.280%#		
21) 2-Butanone-d5	3.873	46	220995	111.147	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery =	111.150%		
24) Chloroform-d	4.349	84	245592	44.340	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	88.680%		
26) 1,2-Dichloroethane-d4	5.031	65	151518	47.557	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	95.120%		
32) Benzene-d6	5.050	84	480456	43.361	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	86.720%		
36) 1,2-Dichloropropane-d6	6.072	67	162802	47.181	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	94.360%		
41) Toluene-d8	7.317	98	421810	41.186	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	82.380%		
43) trans-1,3-Dichloroprop...	7.625	79	67252	39.602	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	79.200%		
47) 2-Hexanone-d5	8.088	63	124717	110.279	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery =	110.280%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	225467	52.728	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	105.460%		
66) 1,2-Dichlorobenzene-d4	11.625	152	195426	48.733	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	97.460%		
Target Compounds						
63) 1,2,4-Trimethylbenzene	10.918	105	18263	1.648	ug/L	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081721\
 Data File : VV021846.D
 Acq On : 18 Aug 2021 04:11
 Operator : SY/MD
 Sample : M3422-22
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKE1

Quant Time: Aug 18 04:38:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 18 03:33:12 2021
 Response via : Initial Calibration

