

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081721\
 Data File : VW021839.D
 Acq On : 18 Aug 2021 01:23
 Operator : SY/MD
 Sample : M3422-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKC8

Quant Time: Aug 18 03:38:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 18 03:33:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	418353	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	386261	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	180490	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	112650	34.444	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.880%
7) Chloroethane-d5	1.558	69	96746	39.200	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.400%
11) 1,1-Dichloroethene-d2	2.105	63	136814	26.280	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.560%#
21) 2-Butanone-d5	3.873	46	227200	111.697	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.700%
24) Chloroform-d	4.349	84	250852	44.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.540%
26) 1,2-Dichloroethane-d4	5.034	65	156374	47.977	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.960%
32) Benzene-d6	5.050	84	493822	44.459	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.920%
36) 1,2-Dichloropropane-d6	6.069	67	165071	47.722	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.440%
41) Toluene-d8	7.317	98	430933	41.975	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.940%
43) trans-1,3-Dichloroprop...	7.625	79	69913	41.069	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.140%
47) 2-Hexanone-d5	8.088	63	120111	105.948	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.950%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	220097	51.347	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.700%
66) 1,2-Dichlorobenzene-d4	11.625	152	187858	50.293	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081721\
Data File : VV021839.D
Acq On : 18 Aug 2021 01:23
Operator : SY/MD
Sample : M3422-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKC8

Quant Time: Aug 18 03:38:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 18 03:33:12 2021
Response via : Initial Calibration

