

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081820\
 Data File : VV018042.D
 Acq On : 18 Aug 2020 10:22
 Operator : SY/MD
 Sample : VV0818MBL01
 Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK83

Quant Time: Aug 19 07:59:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 19 07:54:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	286315	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	277206	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	136752	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	96240	44.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.10%
7) Chloroethane-d5	1.58	69	82677	46.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.52%
11) 1,1-Dichloroethene-d2	2.12	63	144980	34.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.40%
21) 2-Butanone-d5	3.91	46	123773	86.51	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.51%
24) Chloroform-d	4.37	84	175169	44.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.10%
26) 1,2-Dichloroethane-d4	5.05	65	118517	45.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.78%
32) Benzene-d6	5.07	84	352228	46.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.22%
36) 1,2-Dichloropropane-d6	6.09	67	114171	45.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.42%
41) Toluene-d8	7.33	98	312852	45.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.08%
43) trans-1,3-Dichloropropene-	7.64	79	52827	46.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.26%
47) 2-Hexanone-d5	8.11	63	60564	75.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.56%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	134179	41.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.52%
64) 1,2-Dichlorobenzene-d4	11.65	152	129820	48.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.48%

Target Compounds					Ovalue
16) Methylene chloride	2.53	84	2728	1.115 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081820\
Data File : VV018042.D
Acq On : 18 Aug 2020 10:22
Operator : SY/MD
Sample : VV0818MBL01
Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK83

Quant Time: Aug 19 07:59:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 19 07:54:18 2020
Response via : Initial Calibration

