

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV081820\
 Data File : VV018045.D
 Acq On : 18 Aug 2020 11:35
 Operator : SY/MD
 Sample : L3702-21
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Aug 19 08:08:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 19 07:54:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	291391	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	285996	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	138059	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	85932	38.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.30%
7) Chloroethane-d5	1.58	69	76593	42.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.22%
11) 1,1-Dichloroethene-d2	2.12	63	132334	31.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.24%
21) 2-Butanone-d5	3.91	46	119327	81.95	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.95%
24) Chloroform-d	4.38	84	168721	41.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.38%
26) 1,2-Dichloroethane-d4	5.06	65	115473	43.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.86%
32) Benzene-d6	5.08	84	337670	42.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.68%
36) 1,2-Dichloropropane-d6	6.09	67	110609	42.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.84%
41) Toluene-d8	7.34	98	299510	42.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.50%
43) trans-1,3-Dichloropropene-	7.64	79	50059	42.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.66%
47) 2-Hexanone-d5	8.11	63	57337	69.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	69.34%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	131356	39.15	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.30%
64) 1,2-Dichlorobenzene-d4	11.65	152	127169	46.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV081820\
Data File : VV018045.D
Acq On : 18 Aug 2020 11:35
Operator : SY/MD
Sample : L3702-21
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Aug 19 08:08:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 19 07:54:18 2020
Response via : Initial Calibration

