

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081821\
 Data File : VW021861.D
 Acq On : 18 Aug 2021 13:28
 Operator : SY/MD
 Sample : M3422-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKA6

Quant Time: Aug 19 01:17:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 19 01:14:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	367205	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	345312	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	163405	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	144497	50.336	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 100.680%	
7) Chloroethane-d5	1.558	69	113879	52.568	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 105.140%	
11) 1,1-Dichloroethene-d2	2.101	63	151068	33.060	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 66.120%	
21) 2-Butanone-d5	3.873	46	201066	112.618	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 112.620%	
24) Chloroform-d	4.349	84	255612	51.394	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 102.780%	
26) 1,2-Dichloroethane-d4	5.030	65	156618	54.745	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.500%	
32) Benzene-d6	5.050	84	540330	54.414	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 108.820%	
36) 1,2-Dichloropropane-d6	6.069	67	171354	55.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 110.820%	
41) Toluene-d8	7.317	98	491888	53.594	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 107.180%	
43) trans-1,3-Dichloroprop...	7.622	79	74207	48.761	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 97.520%	
47) 2-Hexanone-d5	8.088	63	115006	113.475	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 113.470%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	217315	56.710	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 113.420%	
66) 1,2-Dichlorobenzene-d4	11.625	152	190402	56.303	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 112.600%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081821\
Data File : VV021861.D
Acq On : 18 Aug 2021 13:28
Operator : SY/MD
Sample : M3422-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKA6

Quant Time: Aug 19 01:17:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 19 01:14:55 2021
Response via : Initial Calibration

