

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081821\
 Data File : VW021857.D
 Acq On : 18 Aug 2021 11:50
 Operator : SY/MD
 Sample : M3399-21
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Aug 19 01:16:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 19 01:14:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	414116	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	380724	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	179699	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	147562	45.580	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.160%
7) Chloroethane-d5	1.561	69	115433	47.250	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.500%
11) 1,1-Dichloroethene-d2	2.105	63	179014	34.738	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.480%
21) 2-Butanone-d5	3.876	46	211862	105.222	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.220%
24) Chloroform-d	4.349	84	262452	46.792	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.580%
26) 1,2-Dichloroethane-d4	5.034	65	161847	50.164	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.320%
32) Benzene-d6	5.050	84	554843	50.679	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.360%
36) 1,2-Dichloropropane-d6	6.072	67	172927	50.721	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.440%
41) Toluene-d8	7.317	98	502070	49.615	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.220%
43) trans-1,3-Dichloroprop...	7.625	79	76189	45.407	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.820%
47) 2-Hexanone-d5	8.088	63	113335	101.425	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.420%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	207001	48.994	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.980%
66) 1,2-Dichlorobenzene-d4	11.625	152	191884	51.597	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081821\
Data File : VV021857.D
Acq On : 18 Aug 2021 11:50
Operator : SY/MD
Sample : M3399-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Aug 19 01:16:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 19 01:14:55 2021
Response via : Initial Calibration

