

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW081821\
 Data File : VW021863.D
 Acq On : 18 Aug 2021 14:17
 Operator : SY/MD
 Sample : M3422-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKB1

Quant Time: Aug 19 01:17:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 19 01:14:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	403400	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	372835	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	177851	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	140961	44.698	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.400%
7) Chloroethane-d5	1.558	69	110180	46.297	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.600%
11) 1,1-Dichloroethene-d2	2.101	63	159578	31.789	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.580%
21) 2-Butanone-d5	3.873	46	192500	98.146	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.150%
24) Chloroform-d	4.349	84	251249	45.984	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.960%
26) 1,2-Dichloroethane-d4	5.030	65	152629	48.564	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.120%
32) Benzene-d6	5.050	84	527472	49.198	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.400%
36) 1,2-Dichloropropane-d6	6.072	67	166230	49.788	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.580%
41) Toluene-d8	7.317	98	472791	47.710	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.420%
43) trans-1,3-Dichloroprop...	7.625	79	70545	42.933	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.860%
47) 2-Hexanone-d5	8.088	63	116443	106.411	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.410%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	218132	52.721	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.440%
66) 1,2-Dichlorobenzene-d4	11.625	152	182614	49.614	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081821\
Data File : VV021863.D
Acq On : 18 Aug 2021 14:17
Operator : SY/MD
Sample : M3422-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKB1

Quant Time: Aug 19 01:17:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM081621WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 19 01:14:55 2021
Response via : Initial Calibration

