

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081823\
 Data File : VV031841.D
 Acq On : 18 Aug 2023 13:20
 Operator : SY/MD
 Sample : VV0818WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK354

Quant Time: Aug 19 02:19:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 19 02:19:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	194233	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	209078	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	90428	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	56364	4.878	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.600%
7) Chloroethane-d5	1.532	69	65899	5.439	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.800%
11) 1,1-Dichloroethene-d2	2.056	65	35576	5.330	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.600%
20) 2-Butanone-d5	3.828	46	169309	42.559	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.120%
24) Chloroform-d	4.252	84	172310	5.139	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
26) 1,2-Dichloroethane-d4	4.947	65	84562	5.107	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	4.963	84	278570	4.883	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	5.995	67	100793	4.976	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.249	98	218178	4.201	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	7.561	79	27562	4.433	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.600%
46) 2-Hexanone-d5	8.034	63	128765	43.445	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.880%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	69717	4.963	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.200%
66) 1,2-Dichlorobenzene-d4	11.567	152	85550	5.364	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081823\
Data File : VV031841.D
Acq On : 18 Aug 2023 13:20
Operator : SY/MD
Sample : VV0818WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK354

Quant Time: Aug 19 02:19:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 19 02:19:05 2023
Response via : Initial Calibration

